

Radiation Safety Standard

NOTIFICATION OF APPROVAL OF AN ALTERNATIVE IMAGE QUALITY TEST

I hereby approve the following test for Test 11 (Image Quality), of radiation safety standard HR001:2010 *Standard for radiation apparatus used to carry out plain film diagnostic radiography*.

Test

For extra-oral dental diagnostic radiation apparatus manufactured by Planmeca, the test specified in the document *Planmeca Digital X-ray Unit – Constancy test according to IEC 61223-3-4*, publication number 10008319 Version 2, published by Planmeca.

Criteria for Passing the Test

The quality of the image produced is within the limits specified in Planmeca's published acceptance test protocols for the X-ray unit being tested.

This approval commences on the 27 August 2010.



P M Vevers
Delegate of the Chief Executive

27 August 2010
Date