

**SOA599 Patient Care and Miscellaneous Consumables
Evaluation Form**

Category Number and Category	Sub-Category	Generic Product Description	Supplier name	Supplier Product Description	Brand	Supplier Code	Count HHS Pass	Count HHS Fail	Total counts	Final Pass / Fail	BTS Pass / Fail	BTS Justification for final decision	BTS Additional comments (for example past performance of product and supplier)	BTS Rank 1,2,3
24 Infant feeding equipment	g) Infant feeding equipment: breast pump													
24 Infant feeding equipment	g) Infant feeding equipment: breast pump	Breast pump	Midmed	Ameda Platinum Hospital Grade Pump	Ameda	AM.17805T	3	1	4	Pass	Fail	Did not appear to have 2 phase pumping. Manual change of suction and speed. Started as soon as it was turned on. Unable to program Very loud compared to other pumps. Very large and heavy.		
24 Infant feeding equipment	g) Infant feeding equipment: breast pump	Breast pump	Midmed	Ameda Elite Hospital Grade Pump	Ameda	AM.17808T	3	2	5	Pass	Fail	Did not have 2 phase pumping. Manual pump with turn 2 turn dials for speed and suction. No timer or alarms. Can double and single pump through same unit. Quiet in operation. Light unit.		
24 Infant feeding equipment	g) Infant feeding equipment: breast pump		Midmed	Trolley For Ameda Platinum And Elite Breast Pumps	Ameda	AM.17811	0	0	0	Not Evaluated				

RTI RELEASE