

Guideline

Peripherally inserted central venous catheters (PICC)

1. Purpose

This guideline has been developed as part of the I-Care intervention bundle for the management of intravascular devices (IVDs). This guideline provides recommendations regarding best practice for the use and management of invasive devices based on current evidence for the prevention and control of healthcare associated infection (HAI).

2. Scope

This guideline provides information for all employees, contractors and consultants within the Hospital and Health Services, divisions and commercialised business units within the Queensland public health system.

3. Related documents

Authorising Policy and Standard/s

- [NSQHS Standard 3 – Preventing and Controlling Healthcare Associated Infections](#)

Standards, procedures, guidelines

- [Australian guidelines for the prevention and control of infection in healthcare](#)
- [Guideline for surveillance of healthcare associated infection](#)
- [Hand hygiene guideline](#)

Forms, templates

- [Peripherally inserted central catheter: maintenance – Point of care tool](#)

4. Guideline for peripherally inserted central venous catheters (PICC)

Contents

Key critical points	3
General recommendations	3
Education and competency assessment	3
Hand hygiene.....	4
Surveillance	4
Insertion & management requirements	4
Insertion location.....	4
Catheter types and materials	4
Prophylactic antibiotics.....	5
Catheter site selection.....	5
Maximal barrier precautions	6
Skin preparation: insertion site	6
Catheter fixation.....	7
Dressing type and replacement intervals.....	8
Dressings: skin preparation.....	9

Peripherally inserted central venous catheters (PICC)

Chlorhexidine bathing	9
PICC review	9
In-line filters.....	10
Flushing and locking of PICCs	10
General information.....	10
Flushing of PICCs	11
Locking of PICCs	11
IV admixtures.....	12
Replacement of IV fluids	12
Administration set changes	13
Disconnection of administration sets	13
Medication labelling.....	13
Needleless access ports	14
Blood culture collection for diagnosis of BSI.....	15
Culturing of PICC tips.....	16
Ethanol lock therapy.....	16
Catheter duration and replacement	17
Guide-wire exchanges	18
Removal of PICC	18
References	19
Bibliography.....	24

Key critical points

- Only competent staff (or training staff supervised by competent staff) are to insert Peripherally Inserted Central Venous Catheters (PICC).
- Accurate documentation and record keeping should be maintained to ensure patient safety.
- IVD requirements should be constantly reassessed and any non-essential intravenous devices should be promptly removed.

General recommendations

- The clinician should choose an appropriate Intravascular Device (IVD) – consider catheter type, number of lumens, length, type of therapy, site of insertion, risk of complications including infection, and patient factors.⁽¹⁾
- Only competent staff (or training staff supervised by competent staff) should insert IVDs to minimise infection and other complications.^(1, 2)
- The clinician should explain to the patient (if possible) or parent/guardian the procedure and need for catheterisation.
- Environmental control measures (e.g. pulled curtains, closed door) should be taken to eliminate environmental risk factors for all procedures involving PICCs.⁽²⁾
- All sterile fields should be set up immediately prior to any procedure by the clinician or suitably trained assistant.
 - Trolleys/carts that include all necessary supplies should be dedicated for PICC insertion.^(3, 4)
- Accurate documentation and record keeping should be maintained by the clinician to ensure patient safety, to allow for audits, and to track outbreaks of infection. The documentation should include the date and time of insertion including type of IVD, gauge, length of line on insertion and removal, anatomical site, skin preparation solution used, name of operator, site observations and device removal/replacement details.^(5, 6)

Education and competency assessment

All clinicians involved in the insertion and maintenance of IVDs must ensure that this is within their scope of clinical practice, determined by the individual's credentials, education, training, competence and maintenance of performance at an expected level of safety and quality. The clinician's scope of practice is also dependent upon the capacity and capability of the service in which they are working.^(7, 8)

- All staff involved in the insertion and maintenance of IVDs should complete all competency assessments as required by the healthcare facility. A record of this should be maintained by the facility.^(1, 3, 4, 6, 9-14)
- Simulation training of catheter insertion procedures including infection prevention strategies has been shown to be successful in reducing rates of Central Line Associated Bloodstream Infection (CLABSI).^(3, 14-16)
- A proportion of patients will be responsible for their own catheter care when discharged from hospital in between treatment regimens. It is recommended that patients be provided with theoretical and practical training by a clinician.^(9, 11, 13) This should include step-by-step instructions in text and images, of clinical procedures needed for care, including principles and techniques i.e. hygiene, dressing changes, flushing techniques and manipulation of the

Peripherally inserted central venous catheters (PICC)

catheter.^(5, 12) Where possible, controlled testing of the patient's knowledge as well as their practical execution of the techniques should be undertaken.

Hand hygiene

- It is recommended that healthcare workers perform hand hygiene with an antiseptic-containing soap solution or use an alcohol-based waterless cleanser:
 - before and after palpating catheter insertion sites
 - before and after accessing, repairing, or dressing an intravascular catheter; this includes associated components such as administration sets and access ports.^(1, 3, 4, 11-14, 17-20)
- The use of gloves does not obviate the need for hand hygiene.
- It is recommended that the clinician educate patients and carers about the importance of hand hygiene and ask that they remind all caregivers to clean their hands.⁽²⁾

Surveillance

It is recommended that surveillance be conducted in high-risk patient populations by a facility appointed person to determine healthcare associated (HCA) IVD-related Bloodstream Infection (BSI) rates, monitor trends in rates and assist in identifying lapses in infection control practices.

- A facility-appointed person should:
 - report HCA IVD-related BSIs at least monthly to all stakeholders
 - investigate all clusters of HCA IVD-related BSIs for common cause problems
 - investigate all episodes of HCA IVD-related *Staphylococcus aureus* BSI using an Investigation Checklist e.g. The *Staphylococcus aureus* BSI Checklist available from: <https://www.health.qld.gov.au/publications/clinical-practice/guidelines-procedures/diseases-infection/infection-prevention/icare-bsi-checklist.pdf>.
- The introduction of new products or processes should be monitored to identify any increase or decrease in the occurrence of device associated infection.⁽²⁾

Insertion & management requirements

Insertion location

- PICCs should be inserted in an area where asepsis can be maintained (e.g. dedicated procedure room or radiology suite).⁽²¹⁾
- The catheter tip position should be confirmed by a modified chest x-ray study prior to use.⁽²¹⁾
 - The x-ray should allow visualisation of the catheter along the length of the arm, across the axillary and subclavian veins and into the superior vena cava.
- For potentially difficult/complicated insertions due to patient factors (e.g. impalpable vessels, morbid obesity, pre-existing injury), clinicians should consider using ultrasound guided access for PICC placement.^(1, 14, 21-27)

Catheter types and materials

- Polyurethane or silicone catheters are available in varying sizes, with one or multiple lumens.⁽²⁴⁾
- The minimum necessary number of lumens, connectors and ports should be used.^(1, 6, 10, 13, 22, 28, 29)

Peripherally inserted central venous catheters (PICC)

- If total parenteral nutrition is being administered, clinicians should utilise one lumen exclusively for that use.^(13, 28)
- Clinicians should use the smallest gauge of PICC that will accommodate the prescribed therapy to reduce the risk of phlebitis.⁽²¹⁾
 - Polyurethane is a tougher material, enabling thinner lumen walls and larger internal diameters of the lumina; this significantly increases flow rates and reduces the potential for breakage and rupture of the catheter.⁽⁹⁾ However, polyurethane PICCs have a higher risk of thrombosis.
 - The decision to use polyurethane catheters should be balanced against the higher risk of thrombosis with these catheters compared to silicone catheters e.g. haematological malignancies.
- PICCs are usually very long to accommodate different lengths of insertion. They are frequently trimmed prior to insertion to prevent dislodgement and for easy handling.
 - In 2005, the Therapeutic Goods Administration (TGA) reissued a report related to the inadvertent cutting of the guide-wire while trimming the length of the catheter at the time of insertion. TGA recommended following manufacturer's instructions for use when inserting catheters including instructions for shortening the catheter.⁽³⁰⁾
 - Scissors should not be used by clinicians for trimming of PICCs as this may introduce roughness to the catheter, possibly contributing to thrombosis or infection.⁽³⁰⁾
- The use of antimicrobial (antibiotic or antiseptic) coated or impregnated PICCs may provide some additional clinical benefit; however cost effectiveness arguments have not been resolved.
 - There is no antimicrobial PICC available on the current Queensland public health system Standing Offer Arrangement (SOA).
 - The decision to use these should be based on local factors such as infection rates.

Prophylactic antibiotics

- Prophylactic antibacterial or antifungal agents (oral, intranasal or parenteral) are not recommended for routine use at the time of insertion or during use of a PICC to prevent catheter colonisation or bloodstream infection.^(3, 10, 13, 14, 22, 29)
- Anti-infective/microbial lock prophylaxis - additional studies are required before antimicrobial lock solutions instilled into the catheter lumen(s) can be recommended for preventing BSIs due to concerns of toxicity and emergence of antimicrobial resistance.^(10, 13, 21, 22, 31-33)

Catheter site selection

- Generally PICCs are inserted into the basilic and cephalic veins of the antecubital space^(23, 24) or brachial veins.⁽²²⁾
 - The basilic vein is preferred as it offers the largest diameter of upper extremity vessels and affords a non-tortuous entry into the subclavian vein.
 - The cephalic vein (~6mm) is smaller than the basilic vein (~8mm) and angles 90 degrees to enter the terminal portion of the axillary vein, sometimes making catheter advancement difficult.
 - Brachial veins lie deep in the centre of the mid to upper arm and cannot be outwardly visualised or palpated; ultrasound guidance is required for access.

Peripherally inserted central venous catheters (PICC)

- The most appropriate location for the tip of PICCs is the lower one-third of the superior vena cava (SVC), close to the junction of the SVC and the right atrium.⁽²⁴⁾
- If possible, clinicians should consider the non-dominant arm for ease of self-care.
- PICCs are not advisable in patients with renal failure and impending need for dialysis, in whom preservation of upper-extremity veins is needed for fistula or graft implantation.^(11, 34)

Maximal barrier precautions

Before placing a PICC (including guide-wire exchanges), the operator and any person who enters the sterile field to assist in the procedure should use maximal barrier precautions including a cap, mask, sterile gown, sterile gloves, and a sterile full body drape.^(1, 3-5, 10, 12-14, 19, 21, 22, 28, 29)

- Don protective eyewear and surgical mask (the mask should cover the nose and mouth tightly).
- CVC insertion requires surgical aseptic technique⁽²⁾ and therefore a surgical scrub should be performed prior to the procedure.⁽³⁵⁾
- Aseptically don sterile long-sleeved gown.
- Aseptically don sterile surgical gloves (ensure gloves cover cuff of gown).
- Prep catheter insertion site, allow to dry (refer: [Skin preparation: insertion site](#)).
- Drape the entire body of the patient (while maintaining a sterile field) with a large sterile fenestrated drape leaving only a small opening at the insertion site. The wide arc of the guide-wire and the subsequent need to control its free end require adequate draping well beyond its radius.
- A surgical cap should be used to contain hair that may fall across the operator's face during the procedure.

Skin preparation: insertion site

It is recommended that:

- Hair at the insertion site should only be removed by clinicians (prior to antiseptic application), using clippers (not shaved) to improve adherence of the dressing.^(5, 28)
 - The skin should be physically cleaned (if necessary) by the clinician prior to applying the antiseptic solution and inserting the catheter.
 - Removal of skin lipids (defatting) by the clinician with alcohol, ether or acetone is not recommended.⁽²⁸⁾
 - Use alcohol-containing skin preparatory agents if no contraindication exists. The most effective disinfectant (chlorhexidine or povidone iodine) to combine with alcohol has not been established in the literature (be aware that either agent may be contraindicated e.g. sensitivity, allergy)
 - A solution containing 2% chlorhexidine gluconate (CHG) in ≥ 70% (ethyl or isopropyl) alcohol (alcoholic chlorhexidine) should be used by clinicians for preparation of the insertion site.^(5, 20, 36)
- or
- A solution containing povidone-iodine 10% in 70% ethyl alcohol (ethanol)⁽³⁷⁾ (povidone-iodine should remain on the skin for at least two minutes and until dry before inserting the catheter).
 - Non-sterile antiseptic applicators (e.g. swabsticks) should not be placed on the sterile field. Antiseptic liquid solutions are able to be poured into a sterile pot on the sterile field.⁽²⁸⁾

Peripherally inserted central venous catheters (PICC)

- If using non-sterile antiseptic applicators, skin preparation should be undertaken by an alternative staff member who is not gowned and gloved to insert the line.
- If alcohol is contraindicated (e.g. allergy, sensitivity, skin condition) clinicians should use aqueous povidone-iodine⁽³⁸⁾ 10%* or sterile normal saline 0.9% (*NB: the drying time for aqueous based antiseptics is longer than alcohol based products).
- Note: The same antimicrobial agent shall be used for all phases of the patient's skin preparation, to ensure full residual benefit and consistent action.⁽³⁹⁾
- 70% alcohol solution (including alcohol-impregnated swabs) should not be used as it has no residual antimicrobial activity on the skin.
- The solution should be applied vigorously by the clinician to an area of skin approximately 30cm in diameter, in a circular motion beginning in the centre of the proposed site and moving outward, for at least 30 seconds.⁽⁵⁾
 - This step should be repeated a total of three times using a new swab for each application.
- The clinician should allow the antiseptic to air dry completely prior to inserting the catheter; do not wipe or blot.⁽¹⁾
- Clinicians should not palpate the insertion site after the application of antiseptic, unless aseptic technique is maintained.^(1, 11)
- Clinicians should not routinely use antimicrobial ointments or creams under the dressing at the insertion site.^(1, 10, 12-14, 22, 28, 29)
- The length of the line used should be noted prior to insertion and clearly documented in the patient's notes.⁽⁵⁾

Catheter fixation

- Adhesive tape (alone) should not be used by clinicians to secure PICCs.
- Secure the catheter by:
 - suturing at the hub and three-way bifurcation anchor point, or
 - utilising a sutureless fixation/securement device.
 - A sutureless securement device has been shown to be superior in reducing infection risk,^(11, 14, 40) length of time required to secure the catheter to the skin and avoiding the additional risk of needlestick injury associated with suturing.⁽¹⁾
 - The potential for this device to reduce infection may derive from the elimination of skin suture wounds that are contiguous to the newly inserted catheter and from minimisation of the to-and-fro positioning of the catheter, which may promote invasion of the tract by cutaneous microorganisms through capillary action.^(1, 14, 41)
 - Sutureless securement devices should be used in accordance with manufacturer's instructions.
- A catheter that has migrated externally should not be readvanced prior to re-stabilisation and a chest x-ray should be performed to confirm tip position.⁽⁵⁾

Dressing type and replacement intervals

- PICC dressings should be changed 24 hours after insertion^(6, 40, 42, 43) and as below:

Table 1: Dressing types and replacement intervals

Dressing type	Replacement interval
Transparent, semi-permeable, self-adhesive polyurethane	Weekly ^{*(1, 3, 6, 11, 13, 14, 17, 19, 20, 22, 28, 29, 41, 42)}
Gauze	Second daily ^{*(3, 6, 14, 29, 41, 42)}
Chlorhexidine-impregnated	Weekly ^{*(17)}

**All dressings should be replaced routinely as well as when the dressing becomes damp, loosened, no longer occlusive or adherent, soiled, if there is evidence of inflammation, or excessive accumulation of fluid. Manufacturer's recommendations should be followed.*^(1, 3, 6, 11, 13, 14, 17, 19, 20, 22, 28, 29, 41, 42)

- For longer-term catheter maintenance in home patients, less frequent dressing changes may be possible depending on patient characteristics relating to perspiration and cleanliness. Semi-permeable dressings generally begin to degrade two weeks after application.
- Patient as well as environmental factors should be considered when selecting the most appropriate dressing for use on a peripherally inserted central venous catheter. The following recommendations should be considered:
 - Transparent, semi-permeable, self-adhesive, (standard or hyperpermeable), polyurethane dressings.^(6, 13, 19, 21, 28, 40, 42, 44)
 - Benefits include protecting the site from extrinsic contamination, allowing continuous observation of the insertion site, and helping stabilise and secure the catheter.^(6, 28, 41, 44)
 - The clinician should inspect the dressing on the exit site each shift.⁽²⁰⁾
 - Sterile gauze dressing secured with adhesive tape or semi-permeable dressing.^(19, 42, 44)
 - Gauze dressings should only be used by clinicians if there is a true contraindication to polyurethane dressings including diaphoresis or excessive ooze from the insertion site and should be replaced by a transparent dressing as soon as possible.^(13, 28)
 - If gauze is used in combination with a semi-permeable dressing, it is considered a gauze dressing and should be changed every 48 hours.^(1, 14, 17, 28)
 - Chlorhexidine-impregnated dressings and sponges have been shown to reduce the risk of exit site infection and catheter-related bacteraemia.^(11, 14, 21, 28, 29, 45-47)
 - The safety of chlorhexidine-impregnated dressings/sponges has not been established in low birth weight neonates who may be at risk of skin or systemic toxicity.
- The dressing (including polyurethane types) should not be immersed or submerged in water:⁽¹⁾
 - Showering is preferable to bathing, and swimming or spa bathing should be avoided with any external catheter,⁽²²⁾ in order to prevent colonisation with Gram negative organisms, especially *Pseudomonas* spp.

Peripherally inserted central venous catheters (PICC)

- Clinicians should utilise an aseptic technique^(6, 13, 42) including sterile dressing (or dressing change) pack with drape and sterile gloves when changing the dressing on a PICC.
 - If the patient is coughing or cannot turn their head away from the exit site, consider having them wear a face mask.
- Clinicians should dress each PICC/catheter as a separate procedure.

Dressings: skin preparation

- 2% alcoholic chlorhexidine is the preferred solution for skin preparation for dressings, however if contraindicated, clinicians should use the same solution as for skin preparation prior to PICC insertion (refer: [Skin preparation: insertion site](#)).^(1, 3, 5, 6, 12, 13, 18-20, 22, 28, 42, 48)
- Most PICC and other catheter materials are generally alcohol-resistant, however alcohol can damage some types of polyurethane and silicone PICC tubing (refer to manufacturer's instructions).^(1, 5, 13)
- Removal of skin lipids (defatting) by clinicians with alcohol, ether or acetone is not recommended.⁽²⁸⁾
- Clinicians should remove blood or ooze from catheter insertion site with sterile 0.9% sodium chloride.
- Clinicians should cleanse the area (the size of the final dressing) around the catheter including under the hub.
- Cleansing should be performed using a circular motion moving in concentric circles from the site outward.
 - Clinicians should repeat this step a total of three times using a new swab for each application.⁽¹⁾
- Clinicians should apply the antiseptic solution vigorously for at least 30 seconds and allow to air dry; do not wipe or blot.
- Clinicians should not use antimicrobial ointments or creams under the dressing at the insertion site.^(1, 10, 12, 13, 19, 28)
- Antiseptic-impregnated (chlorhexidine gluconate) dressings/sponges have been shown to be effective in reducing vascular catheter bacterial colonisation and bacteraemia.^(21, 45, 46)
 - The safety of these dressings/sponges has not been established in low birth weight neonates who may be at risk of skin or systemic toxicity.

Chlorhexidine bathing

For information regarding chlorhexidine bathing please refer to appendix two of the [Guideline for the management of multi-resistant organisms](#).

PICC review

- PICCs should be reviewed each shift by clinicians, and those that are no longer clearly needed should be promptly removed.^(1, 4, 11, 22, 29)
 - Insertion site
 - The insertion site should be examined daily by the clinician⁽²⁰⁾ (or at each dressing change if gauze is used) for erythema, exudate, tenderness, pain, redness, swelling, suture integrity and catheter position.^(6, 14)

Peripherally inserted central venous catheters (PICC)

- Signs of systemic infection
 - Site appearance should not be used as the only indicator of infection. Local inflammation is uncommon with PICC-related infection caused by coagulase-negative staphylococci as this pathogen incites little local or systemic inflammation. The patient should also be examined for fever or other signs of sepsis e.g. tachycardia, tachypnoea, hypotension.
- Patency of lumens⁽²⁸⁾
- Patients should be encouraged (where possible) by clinicians to report any changes in their catheter site or any new discomfort.⁽¹⁾

In-line filters

- In-line filters are not recommended for infection control purposes, however certain chemotherapeutic and immunological drugs require filtering for other reasons.⁽⁴³⁾
 - Lines containing filters should be removed by clinicians immediately following administration of the drug.

Flushing and locking of PICCs

General information

- Where possible, continuous intravenous fluids should be administered by clinicians using an infusion pump.
- The optimal volume and frequency of flushing and/or locking of catheters for intermittent injections or infusions is unclear:
 - the literature suggests the volume of the flush or lock should equal at least twice the volume of the catheter plus add-on devices (if used)^(5, 17)
 - the volume of a lumen is generally less than 1mL and a needleless access device 0.1mL, therefore a (minimum) 2-3mL of solution should be sufficient.
- Only single-dose solutions should be used.⁽⁴²⁾
- Clinicians should use a syringe with the internal diameter of a 10mL syringe (or larger), to avoid excessive pressure and catheter rupture (the diameter of 10mL syringes varies slightly between manufacturers but is usually around 14.5-15.5mm). Syringes with an internal diameter smaller than that of a 10mL syringe can produce higher pressure in the lumen and rupture the catheter).^(17, 42)
 - Infusion pressure should never exceed 25 psi because pressures higher than that may also damage blood vessels.
 - The internal diameter of a standard 3mL syringe generates pressure greater than 25 psi, whereas a syringe with the internal diameter of a 10mL syringe generates less than 10 psi.⁽⁶⁾
 - 3mL syringes with the internal diameter of a 10mL syringe do not produce higher pressure and are acceptable for use.
- Clinicians should flush in a pulsatile (push-pause or start-stop-start) motion.^(6, 17, 42, 49)
- Clinicians should use an aseptic technique^(6, 13) including meticulously cleaning the access port(s) with a single-use 70% alcohol-impregnated swab or 2% alcoholic chlorhexidine vigorously for at least 15 seconds and allowing to dry prior to accessing the system.^(3, 13, 17, 42)

Peripherally inserted central venous catheters (PICC)

- Disconnecting the flush syringe allows reflux of blood into the tip of the catheter to displace the space occupied by the syringe. To prevent this source of occlusion clinicians should clamp the extension set or withdraw the syringe while administering the last 0.5mL of flush (positive pressure technique).^(6, 9, 17, 42, 49)
- Positive- or negative-pressure mechanical valve needleless connectors have been associated with increases in rates of catheter-related bacteraemia and therefore are not recommended for use.^(28, 50-52)

Flushing of PICCs

- Flushing is recommended to promote and maintain patency and prevent the mixing of incompatible medications and solutions. Sterile 0.9% sodium chloride for injection should be used by clinicians to flush a catheter unless the manufacturer recommends flushing with heparin sodium solution.⁽¹³⁾
- Clinicians should flush catheters immediately:
 - after placement
 - prior to and after fluid infusion or injection (as an empty fluid container lacks infusion pressure and will allow blood reflux into the catheter lumen from normal venous pressure)
 - prior to and after blood drawing.
- The flush solution and flushing intervals should be documented by the clinician in the patient record.

Locking of PICCs

- PICCs should not be left *in situ* when no longer required as the risk of infection is greater than risks associated with reinsertion of a new device.
- If it is necessary to lock a PICC the following recommendations should be considered by the clinician:
 - Locking involves instilling a solution to prevent occlusion when the device is not in use.
 - There is limited information concerning the most appropriate solution to lock a catheter. Heparinised saline has been used primarily due to the antithrombotic properties of heparin. However, complications such as heparin-induced thrombocytopenia (HIT), altered coagulation studies and bleeding have been reported, particularly if other general anticoagulant therapy is administered.^(18, 53) Additionally, heparin is incompatible with certain substances in solution e.g. gentamicin sulphate (refer MIMS Online available from: <https://www.mimsonline.com.au/Search/Search.aspx>).
 - Until there is further evidence, sterile sodium chloride 0.9% should be used by clinicians to lock a catheter that is no longer required for continuous infusions in preparation for future use; unless the manufacturer recommends catheter lumens be locked with an alternate solution.
 - The most important part of locking the catheter is the mechanical action of the procedure itself, designed to prevent backflow of blood into the catheter tip i.e. 'pulsatile' and 'positive pressure' flushing techniques.^(6, 17, 42, 49)
 - Some PICCs integrate valve technology which restricts blood backflow and air embolism by remaining closed when not in use therefore eliminating the need for heparin flushing to maintain patency.

Peripherally inserted central venous catheters (PICC)

- Low-dose oral warfarin or other systemic anticoagulants should not be prescribed for prophylaxis of catheter occlusion.^(10, 12, 13)

IV admixtures

It is recommended that:

- Clinicians should admix all intravenous fluids using an aseptic technique.⁽¹³⁾
- Clinicians should not use containers of intravenous fluid that have visible turbidity, leaks, cracks or particulate matter, or if the manufacturer's expiration date has passed.
- Clinicians should use single-dose vials for parenteral additives or medications when possible.
- Clinicians should use the recommended needle gauge for injecting additives into infusion bags and/or burettes.⁽⁵⁾

Replacement of IV fluids

Table 2: IV fluid replacement intervals

Fluid	Replacement interval
Standard (crystalloid) and non-lipid parenteral solutions	Every 24 hours
Lipid-containing solutions	Within 24 hours
Lipid emulsions	Within 12 hours
All blood components (excluding factor VIII or IX for continuous infusion)	Within 4 hours
Drug infusions (e.g. heparin, insulin)	Every 24 hours ^(5, 19, 54)

- When any IVD is resited, it is recommended that both the infusion and administration set be replaced by the clinician regardless of when the infusion was initially commenced.⁽²⁸⁾
- IV administration sets should be spiked into IV fluid bags the whole way.⁽⁵⁵⁾
- Each bag of IV fluid should only be spiked once.⁽⁵⁶⁾
- It is recommended that all IV fluids be stored by facilities according to manufacturer's guidelines.
- It is recommended that bags or bottles of intravenous solution should not be used as a common source of supply for multiple patients.⁽²⁾

Administration set changes

- It is recommended that clinicians should ensure all components of the administration system are compatible (this includes burettes), including needleless intravascular devices to minimise leaks and breaks in the system:
 - Add-on equipment should be of luer-lock design.⁽⁵⁾

Table 3: Administration set replacement intervals

Administration set	Replacement interval
Not containing lipids, blood or blood products	Up to 96 hours ^{*(2, 3, 13, 43)}
Lipid/lipid-containing parenteral nutrition	Within 24 hours ^{*(1, 13, 19, 28, 43, 57)}
Chemotherapeutic agents	Remove immediately after use [*]
Propofol	Within 12 hours or as per manufacturer ^{*(1, 2, 43)}
Heparin	Every 24 hours ^(5, 19, 54)
Other infusions (not including blood products)	When disconnected or new catheter ^{*(28)}
[*] All administration sets should be replaced when disconnected or if the catheter is changed ^(1, 19, 28, 57) When an administration set is changed, the IV fluid bag should also be changed. ⁽⁵⁶⁾	

Blood components

- Must be transfused using an administration set approved for this purpose, incorporating a standard filter which removes clots and small clumps of debris that may form during collection and storage. The recommended filter pore size is 170-200 micron.⁽⁵⁾
- Any number of red cell units may be transfused during a 12-hour period provided the flow rate remains adequate. However specific manufacturer's recommendations defining the maximum number of units per blood administration set must not be exceeded.⁽⁵⁴⁾ Administration sets should be removed by the clinician immediately after use.^(13, 20)

Disconnection of administration sets

- Administration sets should not be intermittently disconnected (including for patient showering/toileting).⁽²⁾
- If administration sets are disconnected from the intravascular device, the set should be discarded and a new administration set connected using aseptic technique and observing standard precautions.
- Intermittent disconnection of administration sets increases risk of infection through manipulation of the hub and contamination, and occlusion due to reflux of blood into the catheter tip.^(5, 9, 28)

Medication labelling

- It is recommended that clinicians abide by labelling recommendations for all injectable products prepared in the ward or clinical area, including recommendations for labelling containers (bags, bottles and syringes) and conduits (lines and catheters).^(5, 58)
- It is recommended that Clinicians ensure labelling complies with the national recommendations for user-applied labelling of injectable medicines, fluids and lines (current edition) as set out by [The Australian Commission on Safety and Quality in Healthcare](#).

Needleless access ports

- Clinicians should minimise catheter manipulation (e.g. number of intermittent infusions).⁽⁴³⁾
- Closed catheter access systems are associated with fewer CRBSIs than open systems.⁽¹⁾ Therefore, needleless access ports should be used on all lumens.
 - Stopcocks should be end-capped with a needleless access port/cap when not in use.^(1, 21)
- All persons handling or accessing the intravascular system should first perform hand hygiene.^(20, 22, 28, 29, 42)
- Needleless access ports should be used by clinicians according to manufacturer's recommendations.
- Clinicians should not use adhesive tape as a means of junction securement between the hub and connector or infusion line.
- All intravenous access ports should be meticulously cleaned by the clinician with a single-use 70% alcohol-impregnated swab or 2% alcoholic chlorhexidine vigorously for a minimum of 15 seconds and allowed to dry prior to accessing the system.^(5, 13, 42) For example a typical intermittent infusion of medication may involve swabbing the access port:
 - before the initial saline injection to assess catheter patency,
 - before attaching the sterile infusion tubing or syringe, and
 - before flushing and/or locking the catheter with saline after administering the medication.
- The access port should be accessed by the clinician with a sterile single-use device.
- Anytime an access port is removed from a catheter, the clinician should discard it and a new sterile access port should be attached, using appropriate aseptic technique.
- The integrity of the access port should be confirmed by the clinician before and immediately after each use. If the integrity of the access port is compromised or if residual blood remains within the access port, it should be replaced immediately and consideration given to changing the administration set.⁽⁵⁾
- Needleless access ports should be changed as per manufacturer's instructions, or if the integrity of the port is compromised.⁽⁵⁾ In general, a lot of manufacturers recommend that their needleless components be changed weekly or when there are signs of blood, precipitate, leaks or other defects.⁽⁴²⁾
 - CDC guidelines currently recommend that needleless components be changed at least as frequently as the administration set, but no more frequently than every 72 hours.⁽¹⁾ A recent study has identified an increased CLABSI rate when needleless access ports were changed every 24 hours with lines containing blood products or lipids.⁽⁵⁹⁾
 - More frequent changing of access ports may reduce the burden of access port contamination that could lead to bloodstream infection, however more frequent manipulation of the catheter for access port changes could increase the risk of infection.⁽⁵⁹⁾
- Clinicians can use central venous catheters for blood sampling; it is advisable to follow manufacturers' instructions as to the best practise for their device. With some devices it is advisable to limit or avoid this practice because of the increased risk of occlusion (clotting) and infection in the catheter from any residual blood;^(60, 61) if necessary clinicians should minimise blood sampling by batching laboratory specimen draws.⁽⁴³⁾

Blood culture collection for diagnosis of BSI

Refer to local hospital procedure for blood culture collection and [Pathology Queensland and CHRISP Recommendations for Blood Culture Collection - Adults](#) (Queensland Health Intranet access only).

It is recommended that:

- Blood cultures should always be collected by a clinician from a peripheral vessel:
 - Approximately 20mL is required and 10mL should be placed in each of the anaerobic and aerobic blood culture bottles.⁽⁶¹⁻⁶³⁾
 - Staff should read the instructions on the blood culture bottle as different blood culture systems have different requirements.
 - Each anaerobic and aerobic bottle constitutes a blood culture 'set'. No more than three sets are required in one episode. Two sets has a sensitivity of >90% while collecting three sets will increase that to >98%.⁽⁶⁴⁾
- 10mL draws are suggested for each bottle. There is no need to collect more than two bottles per lumen.⁽⁶¹⁾
- Taking blood cultures through a PICC is discouraged as the practice may cause occlusion and contribute to catheter lumen colonisation.⁽⁶¹⁾
- Blood for culture should only be collected from a PICC in addition to peripheral blood where:
 - there is no other access available, or
 - following placement of a new PICC and only by the operator, or
 - attempting to determine if the catheter (lumen) is contaminated.^(14, 65, 66)
- If catheter-related bloodstream infection is suspected:
 - the clinician should use strict aseptic technique and hand hygiene prior to blood culture collection to reduce the risk of microbial contamination⁽⁶⁷⁾
 - the clinician should utilise sterile collection equipment
 - the clinician should use standard precautions when collecting blood cultures, including eye protection
 - Non-sterile gloves can be used in accordance with aseptic technique. If key parts or key sites are touched, sterile gloves should be used.⁽⁶⁷⁾ If there is a high rate of contamination, routine sterile gloving and/or sterile blood culture kits have been shown to significantly decrease contamination rates.⁽⁶⁸⁻⁷⁰⁾ Cost vs benefit should be considered.^(67, 71)
 - the first sample is to be taken peripherally by the clinician; cleanse skin with alcoholic chlorhexidine or ≥70% alcohol^(67, 72-75) and allow to dry prior to venepuncture
 - additional specimens can be collected by the clinician from each lumen of the catheter as above.^(14, 29, 62, 63)
- The blood culture bottle diaphragm should be swabbed by the clinician with a single-use 70% alcohol-impregnated swab prior to inoculating the bottle.⁽⁶⁷⁾
- There is no need for the clinician to change the blood culture collection needle between venepuncture and bottle inoculation⁽⁶⁴⁾ (careful skin preparation is a more important factor than changing needles in reducing contamination during blood culture collection).⁽⁶²⁾

Peripherally inserted central venous catheters (PICC)

- Catheter discard blood, arterial line blood, intravenous catheter blood, “left over” blood from blood gas or other analyses should not be used by the clinician for blood cultures. If further blood tubes are required for testing, they should be collected after the blood cultures are drawn.^(64, 67)
- The collection site as well as the patient’s clinical and demographic data should be recorded on the request form by the clinician.⁽⁶¹⁾

Culturing of PICC tips

- Routine culture of catheter tips is not recommended,^(29, 63, 65) however periodic sampling could be considered in the context of measuring the effectiveness of interventions. This should only occur in consultation with Infection Prevention and Control and the Microbiology Laboratory.
- Culture of vascular catheter tips may be useful in confirming the source of line related bacteraemia when performed concurrently with peripheral blood cultures. Depending on local laboratory practice, vascular catheter tips are only processed if there is an associated positive blood culture.⁽⁷⁶⁾ Consult with local laboratory.
- If pus is present at the insertion site, clinicians should swab the site prior to cleaning and send for culture.
- If catheter-related sepsis is suspected:
 - the clinician should clean the skin at the skin-catheter junction with alcoholic chlorhexidine and allow the solution to dry prior to catheter removal – this will minimise skin contamination of the catheter tip
 - the clinician should remove the catheter aseptically
 - a segment of the tip of the catheter (optimum length 5cm) should be submitted.^(29, 65) The tip should be aseptically cut from the end of the catheter directly into a sterile specimen container.⁽⁶³⁾ Transport to laboratory as quickly as possible to prevent excessive drying.⁽⁷⁶⁾
- The clinician should ensure the site and type of catheter are noted on the request form as well as the required clinical and demographic data.⁽⁷⁶⁾

Ethanol lock therapy

Antibiotics may be ineffective in the treatment of infected central venous catheters. This is due to the formation of a biofilm on the internal lumen of the catheter. Biofilm prevents antibiotic penetration to the surface of the inner lumen of the catheter despite appropriate antibiotic therapy. Ethanol locks have been proven to be effective in treating catheter infections and prolonging the life of the central venous catheter.^(33, 77, 78)

Recommendations regarding ethanol lock therapy:

- Commencement of ethanol lock therapy should only occur after the patient has been reviewed by the infectious diseases team and following discussion with the treating consultant.
- Ethanol lock therapy **should not** be used if:
 - the patient is unstable
 - the patient has an exit site infection
 - the patient is pregnant or breast feeding
 - the patient has a *Staphylococcus aureus* bacteraemia, known multi-resistant organism present or fungaemia (including candidaemia).

Peripherally inserted central venous catheters (PICC)

- Ethanol lock therapy can be used if:
 - the patient is stable
 - the patient has a catheter-associated bloodstream infection
 - there is no evidence of exit site infection
 - appropriate antibiotic therapy has been initiated
 - the infectious diseases team and treating consultant agree to commence treatment.
- Prescribing recommendations:
 - Ethanol installation volume and withdrawal volumes and sodium chloride 0.9% flushes including the frequency of locks should be ordered by an appropriate clinician on the patient medication chart.
 - The dwell time for an ethanol lock is four hours. The ethanol lock should be repeated daily by clinicians for 4-5 days.
 - The clinician should aspirate the instilled volume at the conclusion of the dwell time and record this in the patient chart.
 - The volume of ethanol to be instilled equals the volume of the lumen plus any connecting tubing. This volume is determined by the PICC type. Refer to the patient chart notes for the manufacturer and serial number of the inserted PICC. Refer to the manufacturer's reference tables for lumen volume.
- Dilution:
 - The clinician should draw up 3.5mL of alcohol 100% (ethanol) and 1.5mL sterile water for injection in a 10mL syringe (makes a total of 5mL of 70%).
 - The clinician should discard excess drug to leave the required volume for the catheter lumen volume.
 - The clinician should flush the PICC pre and post ethanol lock with sodium chloride 0.9%. Post flushing of the line should only occur after the alcohol volume has been withdrawn from the PICC at the conclusion of the four hour dwell time.
- Refer to: [Flushing and Locking of PICCs](#) for correct technique to access line.

Catheter duration and replacement

- Accurate and early diagnosis of infection is essential. Clinicians should monitor for signs of infection including erythema, pain, exudate, heat, tenderness, swelling and systemic signs of sepsis.
- The PICC should only be replaced on clinical indications, i.e. clinical infection +/- purulence at the insertion site.^(1, 10, 14)
- PICCs should not be routinely replaced.^(13, 14)
- Because breaches in sterile technique are more likely during emergency procedures, PICCs inserted during a medical emergency should be replaced as soon as possible^(2, 28) and no longer than 48 hours.
- Patients transferring from other healthcare facilities with a PICC *in situ* should have this device reviewed upon arrival by a clinician for infectious and mechanical complications.

Peripherally inserted central venous catheters (PICC)

- PICCs should be reviewed each shift by clinicians, and those that are no longer clearly needed should be promptly removed.^(1, 3, 11, 12, 20, 22, 28, 29)
- Clinicians should replace all fluid administration tubing and connectors when the PICC is replaced.^(1, 19, 57)

Guide-wire exchanges

- Clinicians should not use guide-wire exchanges routinely for percutaneous catheters to prevent infection.^(1, 3, 13, 14, 28) The exception may be early failure of the device in a situation where a new central venous puncture would be hazardous to the patient.
- Guide-wire exchanges of temporary PICCs should not occur in the presence of BSI.
- For guide-wire exchanges, clinicians should use the same meticulous aseptic technique and use of full sterile barriers as used during the insertion of any new PICC.
- After vigorously cleansing the site with the antiseptic solution, inserting the guide-wire, removing the old catheter, and cleaning the site once more with the antiseptic solution, the operator should re-glove and re-drape the site, as the original gloves and drapes are likely to have become contaminated from manipulation of the old catheter.^(1, 28)

Removal of PICC

Also refer to local hospital procedure for removal of PICC.

- The clinician should:
 - perform hand hygiene and don non-sterile gloves
 - remove dressing and securement device
 - clean the site thoroughly with alcoholic chlorhexidine and allow to dry prior to removal of catheter.
- Simple traction by the clinician can remove the catheter.⁽⁷⁹⁾
- Digital pressure should be applied by the clinician until haemostasis is achieved.^(5, 79)
- The clinician should cover the site with gauze and a transparent dressing; the dressing should be changed and the access site assessed every 24 hours until the site is epithelialized.
- On removal the clinician should visually check the integrity of the line to ensure that the tip is present, the complete line has been removed and no breakage has occurred.
- The removed line should be measured and its length documented and checked against the length documented on insertion.

References

1. O'Grady N, Alexander M, Burns L, Dellinger E, Garland J, Heard S, et al. Healthcare Infection Control Practices Advisory Committee (HICPAC). Guidelines for the prevention of intravascular catheter-related infections. 2011.
2. ACSQHC. Australian Guidelines for the Prevention and Control of Infection in Healthcare. National Health and Medical Research Council; 2010.
3. Marschall J, Mermel L, Fakih M, Hadaway L, Kallen A, O'Grady N, et al. Strategies to prevent central line-associated bloodstream infections in acute care hospitals: 2014 update. *Infection Control and Hospital Epidemiology*. 2014;35(7):753-71.
4. Sacks G, Diggs B, Hadjizacharia P, Green D, Salim A, Malinoski D. Reducing the rate of catheter-associated bloodstream infections in a surgical intensive care unit using the Institute for Healthcare Improvement central line bundle. *The American Journal of Surgery*. 2014;207(6):817-23.
5. Dougherty L, Bravery K, Gabriel J, Kayley J, Malster M, Scales K, et al. Standards for infusion therapy (third edition). Royal College of Nursing; 2010.
6. Ho A, Bravery K. Central venous access devices (long term). London: Great Ormond Street Hospital for Children NHS Foundation Trust. 2013.
7. Queensland Government. Health service directive # QH-HSD-034:2014 Credentialing and defining the scope of clinical practice. 2014.
8. ACSQHC. Standard for credentialing and defining the scope of clinical practice. Australian Commission on Safety and Quality in Health Care. 2004.
9. Adams S, Barrett L, Brooks S, Dahler A, Jansens W, Shaw H. Central venous access devices: principles for nursing practice and education. Cancer Nurses Society of Australia; 2007.
10. Schiffer C, Mangu P, Wade J, Camp-Sorrell D, Cope D, El-Rayes B, et al. Central venous catheter care for the patient with cancer: American Society of Clinical Oncology Clinical Practice Guideline. *Journal of Clinical Oncology*. 2013;31(10):1357-70.
11. Bowe-Geddes L. Planning for and successfully managing long-term venous access devices. *Mosby's Nursing Consult*. 2013 Dec 20, 2013.
12. Camp-Sorrell D. State of the science of oncology vascular access devices. *Seminars in Oncology Nursing*. 2010;26(2):80-7.
13. Loveday H, Wilson J, Pratt R, Golsorkhi M, Tingle A, Bak A, et al. epic3: national evidence-based guidelines for preventing healthcare-associated infections in NHS hospitals in England. *Journal of Hospital Infection*. 2014(86S1):S1-S70.
14. Hentrich M, Schalk E, Schmidt-Hieber M, Chaberny I, Mousset S, Buchheidt D, et al. Central venous catheter-related infections in hematology and oncology: 2012 updated guidelines on diagnosis, management and prevention by the Infectious Diseases Working Party of the German Society of Hematology and Medical Oncology. *Annals of Oncology*. 2014;25(5):936-47.
15. Allen G, Miller V, Nicholas C, Hess S, Cordes M, Fortune J, et al. A multitiered strategy of simulation training, kit consolidation, and electronic documentation is associated with a reduction in central line-associated bloodstream infections. *American Journal of Infection Control*. 2014;42:643-8.

Peripherally inserted central venous catheters (PICC)

16. Barsuk J, Cohen E, Potts S, Demo H, Gupta S, Feinglass J, et al. Dissemination of a simulation-based mastery learning intervention reduces central-line associated bloodstream infections. *BMJ Quality and Safety*. 2014;23:749-56.
17. Nailon R, O'Neill S, Cowdery P, Wardian Hartung S, Tyner K, Tomb P, et al. Standardizing central venous catheter care: hospital to home. Rockville MD: Agency for Healthcare Research and Quality (AHRQ); 2012.
18. Bhola C, Lok C. Central venous catheters: optimizing the suboptimal. *Nephrology News and Issues*. 2011 April 27, 2011.
19. Zingg W, Cartier-Fassler V, Walder B. Central venous catheter-associated infections. *Best Practice and Research: Clinical Anaesthesiology*. 2008;22(3):407-21.
20. McIntyre H, Raw A, Stephenson J, Bradley A, Dawson J, Fay M, et al. NICE quality standard 61: Infection prevention and control. National Institute for Health and Care Excellence; 2014.
21. Rupp S, Apfelbaum J, Blitt C, Caplan R, Connis R, Domino K, et al. Practice guidelines for central venous access: a report by the American Society of Anesthesiologists Task Force on Central Venous Access. *Anesthesiology*. 2012;116(3):539-73.
22. Smith R, Nolan J. Central venous catheters. *BMJ*. 2013;347:28-32.
23. Gong P, Huang X, Chen C, Liu J, Meng A, Feng J. Comparison of complications of peripherally inserted central catheters with ultrasound guidance or conventional methods in cancer patients. *Asian Pacific Journal of Cancer Prevention*. 2012;13(5):1873-5.
24. Johansson E, Hammarskjold F, Lundberg D, Arnlind M. Advantages and disadvantages of peripherally inserted central venous catheters (PICC) compared to other central venous lines: a systematic review of the literature. *Acta Oncologica*. 2013;52:886-92.
25. Association for Vascular Access. The use of ultrasound guidance by registered nurses for central venous catheter insertion (position statement). 2010 17 June.
26. Association for Vascular Access. The use of seldinger or modified seldinger technique, in combination with real-time imaging modalities for peripherally inserted central catheter and midline placements by clinicians (position statement). 2011 20 June.
27. Cartier V, Haenny A, Inan C, Walder B, Zingg W. No association between ultrasound-guided insertion of central venous catheters and bloodstream infection: a prospective observational study. *Journal of Hospital Infection*. 2014;87:103-8.
28. Australian and New Zealand Intensive Care Society (ANZICS). Central line insertion and maintenance guideline. Australian Commission on Safety and Quality in Health Care. 2012 April 2012.
29. Cameron J, Cameron A. Central line-associated bloodstream infections. In: Cameron C, Cameron A, editors. *Current Surgical Therapy*, eleventh edition. Philadelphia, PA: Saunders; 2014.
30. Therapeutic Goods Administration. Potential embolisation and/or unravelling of guide-wires following insertion of central venous catheters. Department of Health, Australian Government. 2005.
31. van de Wetering M, van Woensel J, Lawrie T. Prophylactic antibiotics for preventing gram positive infections associated with long-term central venous catheters in oncology patients (review). *The Cochrane Library*. 2013(11).

Peripherally inserted central venous catheters (PICC)

32. Natarajan S, Narayan S, Liu Yin J. Update on the management of febrile neutropenia. *European Oncology & Haematology*. 2011;7(4):263-9.
33. Chaudhury A, Rangineni J, Venkatramana B. Catheter lock technique: in vitro efficacy of ethanol for eradication of methicillin-resistant staphylococcal biofilm compared with other agents. *Federation of European Microbiological Societies: Immunology & Medical Microbiology*. 2012;65:305-8.
34. Association for Vascular Access. Preservation of peripheral veins in patients with chronic kidney disease (position statement). 2011 14 March.
35. Grayson L, Russo P, K R, S H, K H. Five moments for hand hygiene. *Hand Hygiene Australia*. 2013.
36. NSW Ministry of Health. Guideline for Peripheral Intravenous Cannula (PIVC) insertion and post insertion care in adult patients. 2013.
37. Morris W, Tay M. Strategies for preventing peripheral intravenous cannula infection. *British Journal of Nursing*. 2008;17(19):S14-S21.
38. American Society of Anesthesiologists Task Force (ASA). Practice guidelines for central venous access. *Anesthesiology*. 2012;116(3):539-73.
39. Australian College of Operating Room Nurses (ACORN). 2014-2015. Standards for perioperative nursing.
40. Weinstein S, Hagle M. *Plumer's principles and practice of infusion therapy* (9th edition). Philadelphia, PA: Lippincott Williams & Wilkins; 2014.
41. Petree C, Wright D, Sanders V, Killion J. Reducing blood stream infections during catheter insertion. *Radiologic Technology*. 2012;83(6):532-40.
42. Centers for Disease Control and Prevention (CDC). Basic infection control and prevention plan for outpatient oncology settings. Atlanta, GA: Division of Healthcare Quality Promotion (DHQP); 2011.
43. Davis M. Pediatric central venous catheter management: a review of current practice. *Journal of the Association for Vascular Access*. 2013;18(2):93-8.
44. Webster J, Gillies D, O'Riordan E, Sherriff K, Rickard C. Gauze and tape and transparent polyurethane dressings for central venous catheters (review). *The Cochrane Library*. 2011(11).
45. Scheithauer S, Lewalter K, Schroder J, Koch A, Hafner H, Krizanovic V, et al. Reduction of central venous line-associated bloodstream infection rates by using a chlorhexidine-containing dressing. *Infection*. 2014;42:155-9.
46. Timsit J, Mimos O, Mourvillier B, Souweine B, Garrouste-Orgeas M, Alfandari S, et al. Randomized controlled trial of chlorhexidine dressing and highly adhesive dressing for preventing catheter-related infections in critically ill adults. *American Journal of Respiratory and Critical Care Medicine*. 2012;186(12):1272-8.
47. Safdar N, O'Horo J, Ghufuran A, Bearden A, Didier M, Chateau D, et al. Chlorhexidine-impregnated dressing for prevention of catheter-related bloodstream infection: a meta-analysis. *Critical Care Medicine*. 2014;42(7):1703-13.
48. Bilir A, Yelken B, Erkan A. Chlorhexidine, octenidine or providone iodine for catheter related infections: a randomized controlled trial. *Journal of Research in Medical Sciences*. 2013;18(6):510-2.

Peripherally inserted central venous catheters (PICC)

49. Chong L, Chow Y, Kong S, Ang E. Maintenance of patency of central venous access devices by registered nurses in an acute ambulatory setting: an evidence utilisation project. *International Journal of Evidence Based Healthcare*. 2013;11(1):20-5.
50. Jarvis W. Choosing the best design for intravenous needleless connectors to prevent healthcare-associated bloodstream infections. *Infection Control Today*. 2010;14(8).
51. Jarvis W, Murphy C, Hall K, Fogle P, Karchmer T, Harrington G, et al. Health care-associated bloodstream infections associated with negative- or positive-pressure or displacement mechanical valve needleless connectors. *Clinical Infectious Diseases*. 2009;49:1821-7.
52. Btaiche I, Kovacevich D, Khalidi N, Papke L. The effects of needleless connectors on catheter-related bloodstream infections. *American Journal of Infection Control*. 2011;39(4):277-83.
53. Sona C, Prentice D, Schallom L. National survey of central venous catheter flushing in the intensive care unit. *American Association of Critical-Care Nurses*. 2012;32(1):e12-e9.
54. Australian and New Zealand Society of Blood Transfusion Ltd. Guidelines for the administration of blood products, 2nd edition. Royal College of Nursing, Australia; 2011.
55. Baxter Healthcare Pty Ltd. Correct intravenous practices. Email to Communicable Diseases and Infection Management (CDIM_Infection_Management@health.qld.gov.au) 2014 Dec 1.
56. O'Connell D. Findings of inquest into the death of Ruby Yan Chen. Office of the State Coroner, Queensland Courts. 2014.
57. Ullman A, Cooke M, Gillies D, Marsh N, Daud A, McGrail M, et al. Optimal timing for intravascular administration set replacement (review). *The Cochrane Library*. 2013(9).
58. ACSQHC. National recommendations for user-applied labelling of injectable medicines, fluids and lines. Australian Commission on Safety and Quality in Health Care; 2012.
59. Sandora T, Graham D, Conway M, Dodson B, Potter-Bynoe G, Margossian S. Impact of needleless connector change frequency on central line-associated bloodstream infection rate. *American Journal of Infection Control*. 2014;42:485-9.
60. National Kidney Foundation. KDOQI clinical practice guidelines and clinical practice recommendations for 2006 updates: haemodialysis adequacy, peritoneal dialysis adequacy and vascular access. *American Journal of Kidney Disease*. 2006;48(S1-S322).
61. Halm M, Hickson T, Stein D, Tanner M, VandeGraaf S. Blood cultures and central catheters: is the "easiest way" best practice? *American Journal of Critical Care*. 2011;20:335-8.
62. Pfenninger J, Fowler G. Drawing blood cultures. In: Pfenninger J, Fowler G, editors. *Pfenninger and Fowler's Procedures for Primary Care*, third edition. Philadelphia, PA: Elsevier Mosby; 2011.
63. Septimus E. Clinician guide for collecting cultures. Centers for Disease Control and Prevention. 2014.
64. Heney C. Pathology Queensland/CHRISP recommendations for blood culture collections - adults. Queensland Health; 2014.
65. Mermel L, Allon M, Bouza E, Craven D, Flynn P, O'Grady N, et al. Clinical practice guidelines for the diagnosis and management of intravascular catheter-related infection: 2009 update by the Infectious Diseases Society of America. *Clinical Infectious Diseases*. 2009;49:1-45.

66. Schoot R, Dalen E, Ommen C, van de Wetering M. Antibiotic and other lock treatments for tunneled central venous catheter-related infections, in children with cancer (review). The Cochrane Library. 2013(6).
67. Health Protection Scotland. Targeted literature review: what are the key infection prevention and control recommendations to inform a prevention of blood culture contamination quality improvement tool? Version 2.0: September 2014.
68. Kim N, Kim M, Lee S, Yun N, Kim K, Park S, et al. Effect of routine sterile gloving on contamination rates in blood culture: a cluster randomized trial. *Annals of Internal Medicine*. 2011;154(3):145-51.
69. Self W, Speroff T, Grijalva C, McNaughton C, Ashburn J, Liu D, et al. Reducing blood culture contamination in the emergency department: an interrupted time series quality improvement study. *Academic Emergency Medicine*. 2013;20:89-97.
70. Self W, Mickanin J, Grijalva C, Grant F, Henderson M, Corley G, et al. Reducing blood culture contamination in community hospital emergency departments: a multicenter evaluation of a quality improvement intervention. *Academic Emergency Medicine*. 2014;21(3):274-82.
71. Dawson S. Blood culture contaminants. *Journal of Hospital Infection*. 2014;87:1-10.
72. Caldeira D, David C, Sampaio C. Skin antiseptics in venous puncture-site disinfection for prevention of blood culture contamination: systematic review with meta-analysis. *Journal of Hospital Infection*. 2011;77:223-32.
73. Maiwald M, Chan E. The forgotten role of alcohol: a systematic review and meta-analysis of the clinical efficacy and perceived role of chlorhexidine in skin antisepsis. *PLoS One*. 2012;7(9).
74. Maiwald M, Chan E. Pitfalls in evidence assessment: the case of chlorhexidine and alcohol in skin antisepsis. *Journal of Antimicrobial Chemotherapy*. 2014;69(8):2017-21.
75. Washer L, Chenoweth C, Kim H, Rogers M, Malani A, Riddell J, et al. Blood culture contamination: a randomized trial evaluating the comparative effectiveness of 3 skin antiseptic interventions. *Infection Control and Hospital Epidemiology*. 2013;34(1):15-21.
76. Harper J. Culture of tips and related devices (Pathology Queensland). Queensland Health; 2014.
77. Rajpurkar M, McGrath E, Joyce J, Boldt-MacDonald K, Chitlur M, Lusher J. Therapeutic and prophylactic ethanol lock therapy in patients with bleeding disorders. *Haemophilia*. 2014;20:52-7.
78. Tan M, Lau J, Guglielmo B. Ethanol locks in the prevention and treatment of catheter-related bloodstream infections. *Annals of Pharmacotherapy*. 2014;48(5):607-15.
79. Kaufman J. Chapter 7: Upper extremity, neck and central thoracic veins. In: Kaufman J, Lee M, editors. *Vascular and interventional radiology* (2nd edition). Philadelphia, PA: Elsevier Saunders; 2014.
80. FDA Drug Safety communication: FDA requests label changes and single-use packaging for some over-the-counter topical antiseptic products to decrease risk of infection. U.S. Food and Drug Administration; 2013. Available from <http://www.fda.gov/downloads/Drugs/DrugSafety/UCM374870.pdf> .

Bibliography

1. Australia. Australian Council for Safety and Quality in Health Care. National Strategy to Address Health Care Associated Infection: 2003.
2. Bernath V. Tourniquets in phlebotomy. Australia: Monash University: 2006.
3. Australia. NSW Health. Cancer Institute NSW CI-SCaT. Central Venous Access Devices: 2005.
4. Australia. Queensland Health. Queensland Health Pathology and Scientific Services. 2005 Pathology Manual: 2005.
5. Australia. The Canberra Hospital. Lever D, Dahler A, Boland M. Central Venous Catheter Manual: 2005.
6. Australia. The Royal College of Pathologists of Australasia (RCPA). The RCPA Manual Version 4.0: 2004. Available from: <http://www.rcpamanual.edu.au>
7. Canada. Registered Nurses Association of Ontario. Nursing Best Practice Guideline. Care and Maintenance to Reduce Vascular Access Complications: April 2005.
8. United States. Centers for Disease Control and Prevention. Guidelines for the Prevention of Intravascular Catheter-Related Infections: 2011.
9. Crnich C, Maki D. The Promise of Novel Technology for the Prevention of Intravascular Device-Related Bloodstream Infection. I. Pathogenesis and Short-Term Devices. *Healthcare Epidemiology* 2002; 34: 1232-1242.
10. Crnich C, Maki D. The Promise of Novel Technology for the Prevention of Intravascular Device-Related Bloodstream Infection. II. Long-Term Devices. *Healthcare Epidemiology* 2002; 34: 1362-1368.
11. Crowley J. Vascular Access. *Techniques in Vascular and Interventional Radiology* 2003; 6(4): 176-181.
12. Dudrick S. History of Vascular Access. *Journal of Parenteral and Enteral Nutrition* 2006; 30(1): S47-S56.
13. Ewenstein B, Valentino L, Journeycake J, Tarantino M, Shapiro A, Blanchette S, Hoots W, Buchanan G, Manco-Johnson M, Rivard G, Miller L, Geraghty S, Maahs J, Stuart R, Dunham T, Navickis R. Consensus recommendations for the use of central venous devices in haemophilia. *Haemophilia* 2004; 10(5): 629-648.
14. Fellowes C, Kerstein R, Clark J, Azadian B. MRSA on tourniquets and keyboards. *Journal of Hospital Infection* 2006; 64(1): 87-88.
15. Forseter G, Joline C, Wormser G. Blood contamination of tourniquets used in routine phlebotomy. *American Journal of Infection Control* 1990; 18(8): 389-390.
16. Frey A, Schears G. Why Are We Stuck on Tape and Suture?: A Review of Catheter Securement Devices. *Journal of Infusion Nursing* 2006; 29(1): 34-38.
17. Galloway S, Bodenham A. Long-term central venous access. *British Journal of Anaesthesia* 2004; 92(5): 722-734.
18. Ganeshan A, Warakaulle D, Uberoi R. Central Venous Access. *Cardio Vascular and Interventional Radiology* 2007; 30: 26-33.

Peripherally inserted central venous catheters (PICC)

19. Gillies D, O’Riordan L, Wallen M, Morrison A, Rankin K, Nagy S. Optimal timing for intravenous administration set replacement. *The Cochrane Database of Systematic Reviews* 2005, Issue 4. Art. No.: CD003588.pub2. DOI: 10.1002/14651858.CD003588.pub2.
20. Gillies D, O’Riordan L, Wallen M, Rankin K, Morrison A & Nagy S. Timing Of Intravenous Administration Set Changes: A Systematic Review. *Infection Control and Hospital Epidemiology* 2004; 25(3): 240-250.
21. Grant J. Anatomy and Physiology of Venous System Vascular Access: Implications. *Journal of Parenteral and Enteral Nutrition* 2006; 30(1): S7-S12.
22. Guidelines on Paediatric Parenteral Nutrition. Chapter 9. Venous Access. *Journal of Pediatric Gastroenterology and Nutrition* 2005; 41: S54-62.
23. Hadaway L. Prevent occlusions with these flushing pointers. *Nursing* 2003; 33(1): 28.
24. Infusion Nurses Standards of Practice. *Journal of Infusion Nursing* 2006; 29(1) (Supplement): S1-S62.
25. Isaacman D, Karasic R. Lack of effect of changing needles on contamination of blood cultures. *The Pediatric Infectious Disease Journal* 1990; 9(4): 274-278.
26. Lyon S. Vascular Access Devices and the Oncology Patient. *Cancer Forum* 2005; 29(3): 140-144.
27. Moureau N. Vascular Access Devices. *Nursing* 2001; 31(7): 52-55.
28. Playford E, Looke D, Whitby M, Stackelroth J, Harrison K, Watts A. Endemic Nosocomial Gram-negative Bacteraemias Resulting from Contamination of Intravenous Heparin Infusions. *Journal of Hospital Infection* 1999; 42(1): 21-26.
29. Pratt RJ, Pellowe C, Loveday HP, Robinson N, Smith G, Barrett S, Davey P, Loveday C, McDougall C, Mulhall A, Privett S, Smales C, Taylor L, Weller B, Wilcox MI. The epic Project: Developing National Evidence-based Guidelines for Preventing Healthcare Associated Infections. Guidelines for preventing infections associated with the insertion and maintenance of central venous catheters. *Journal of Hospital Infection*, 2001; 47 (Supplement): S47-S67.
30. Pratt R, Pellowe C, Wilson J, Loveday H, Harper P, Jones S, McDougall C, Wilcox M. epic2 National Evidence-Based Guidelines for Preventing Healthcare-Associated Infections in NHS Hospitals in England. *Journal of Hospital Infection* 2007; 65S: S1-S64.
31. Rickard C, Lipman J, Courtney M, Siverson R, Daley P. Routine Changing Of Intravenous Administration Sets Does Not Reduce Colonization Or Infection In Central Venous Catheters. *Infection Control and Hospital Epidemiology* 2004; 25(8): 650-655.
32. Rosenthal K. Pinpointing intravascular device infections. *Nursing Management* 2003; 34(6): 35-43.
33. Rosenthal K. Optimal infusion therapy? Overcome occlusions. *Nursing Management* 2002; 33(2): 49-50.
34. Rourke C, Bates C, Read R. Poor hospital infection control practice in venepuncture and the use of tourniquets. *Journal of Hospital Infection* 2001; 49: 59-61.
35. Rubinson L, Wu A, Haponik E, Diette G. Why Is It That Internists Do Not Follow Guidelines For Preventing Intravascular Catheter Infections? *Infection Control and Hospital Epidemiology* 2005; 26(6): 525-533.

Peripherally inserted central venous catheters (PICC)

36. Sacar S, Turgut H, Kaleli I, Cevahir N, Asan, Sacar M, Tekin K. Poor hospital infection control practice in hand hygiene, glove utilization, and usage of tourniquets. *American Journal of Infection Control* 2006; 34(9): 606-609.
37. Safdar N, Maki D. Inflammation at the insertion site is not predictive of catheter-related bloodstream infection with short-term, noncuffed central venous catheters. *Critical Care Medicine* 2002; 30(12): 2632-2635.
38. Sansivero G. Venous Anatomy and Physiology: Considerations for Vascular Access Device Placement and Function. *Journal of Infusion Nursing* 1998; 21(5S): S107-S114.
39. United States. Infection Control Today. Hanchett M. Needleless Connectors and Bacteremia: Is There a Relationship? [online] 1 November 2005. Available from: http://www.infectioncontrolday.com/articles/410/410_5b1feat2.html (Internet Access Required)
40. United States. Infection Control Today. Schmidt M. Preventing Intravenous Catheter-Associated Infections: An Update: 2001.
41. United Kingdom. National Patient Safety Agency. Learning through action to reduce infection: 2006.
42. United States. eMedicine. Larson S, Hebra A. Vascular Access: A Surgical Perspective: 2006. Available from: <http://www.emedicine.com/ped/topic3050.htm> (Internet Access Required)
43. United States. Institute for Healthcare Improvement (IHI). Getting Started Kit: Prevent Central Line Infections. How-to Guide: 2006. Available from: <http://www.ihl.org/IHI/Programs/Campaign/> (Internet Access Required)
44. Waitt C, Waitt P, Pirmohamed M. Intravenous therapy. *Postgrad Med J* 2004; 80: 1-6.
45. Australia. Therapeutic Goods Administration. TGA News Issue 46: 2005.
46. Burns D. The Vanderbilt PICC Service: Program, Procedural, and Patient Outcomes Successes. *Journal of the Association for Vascular Access* 2005; 10(4): 1-10
47. Camara D. Minimizing Risks Associated with Peripherally Inserted Central Catheters in the NICU. *The American Journal of Maternal/Child Nursing* 2001; 26(1): 17-22.
48. Carlson K. Correct Utilization and Management of Peripherally Inserted Central Catheters and Midline Catheters in the Alternate Care Setting. *Journal of Infusion Nursing* 1999; 22(S6): 46.
49. Cheong K, Perry D, Karapetis C, Koczwara B. High rate of complications associated with peripherally inserted central venous catheters in patients with solid tumours. *Internal Medicine Journal* 2004; 34: 234-238.
50. Gamulka B, Mendoza C, Connolly B. Evaluation of a Unique, Nurse-Inserted, Peripherally Inserted Central Catheter Program. *Pediatrics* 2005; 115(6): 1602-1606.
51. Knue M, Doellman D, Jacobs B. Peripherally Inserted Central Catheters in Children: A Survey of Practice Patterns. *Journal of Infusion Nursing* 2006; 29(1): 28-33.
52. Kumar M, Amin M. The peripherally inserted central venous catheter; friend or foe? *International Journal of Oral and Maxillofacial Surgery* 2004; 33: 201-204.
53. Massola J, Schott-Baer D, Addy L. Clinical Factors Associated with the Development of Phlebitis After Insertion of a Peripherally Inserted Central Catheter. *Journal of Infusion Nursing* 1999; 22(1): 36-42.

Peripherally inserted central venous catheters (PICC)

54. Moureau N. It's all about PICCs. *Nursing Management* 2006; 37(5): 22-27.
55. Ngo A, Murphy S. A Theory-based Intervention to Improve Nurses' Knowledge, Self-efficacy, and Skills to Reduce PICC Occlusion. *Journal of Infusion Nursing* 2005; 28(3): 173-181.
56. Parvez B, Parmar N. Trimming of peripherally inserted central venous catheters may increase the risk of thrombosis (Letter to the Editor). *Thrombosis Research* 2004; 113: 175-177.
57. Safdar N, Maki D. Risk of Catheter-Related Bloodstream Infection With Peripherally Inserted Central Venous Catheters Used in Hospitalized Patients. *Chest* 2005; 128(2): 489-495.
58. Yamamoto A, Solomon J, Soulen M, Tamg J, Parkinson K, Lin R, Schears G. Sutureless Securement Device Reduces Complications of Peripherally Inserted Central Venous Catheters. *Journal of Vascular and Interventional Radiology* 2002; 13: 77-81.
59. Todd J. Choice and use of peripherally inserted central catheters by nurses. *Professional Nurse* 2004; 19(9).
60. United States. National Association of Vascular Access Networks (NAVAN). NAVAN Position Statement: Tip Location of Peripherally Inserted Central Catheters. Reprinted from *Journal of Vascular Access Devices* 1998; 9: 1-4. Available from: <http://www.avainfo.org/website/download.asp?id=1447> (Internet Access Required)
61. Yap Y, Karapetis C, Lerosé S, Iyer S, Koczwara B. Reducing the risk of peripherally inserted central catheter line complications in the oncology setting. *European Journal of Cancer Care* 2006; 15: 342-347.

5. Definitions of terms used in the policy and supporting documents

Term	Definition / Explanation / Details	Source
Central line associated blood stream infection (CLABSI) / Catheter-related bacteraemia	Blood cultures are positive for the presence of bacteria with or without the accompanying symptom of fever, and no apparent source for the infection other than the catheter.	NKF K/DOQI, 2006 ⁽⁶⁰⁾
Exit-site infection	Inflammation (erythema, warmth, tenderness, induration within 2cm of the exit site) or purulence, confined to the area surrounding the catheter exit site, not extending superiorly beyond the cuff if the catheter is tunneled, with exudate confirmed to be positive by microscopy/culture and no systemic symptoms or positive blood cultures.	NKF K/DOQI, 2006 ⁽⁶⁰⁾
Healthcare Associated Infection (HAI)	Healthcare associated infections (HAI) are those infections that are not present or incubating at the time of admission to a healthcare program or facility, develop within a healthcare organisation or are produced by micro-organisms acquired during admission.	ACSQHC ⁽²⁾
Non-sterile antiseptic applicators (e.g. swabsticks)	Topical antiseptic applicators containing antiseptic solution (e.g. isopropyl alcohol, chlorhexidine gluconate). These applicators are single use only and are not usually manufactured as sterile products. Refer to manufacturer's recommendations and labelling.	U.S. Food and Drug Administration ⁽⁸⁰⁾

6. Approval and Implementation

Document custodian

Dr Heidi Carroll, Senior Medical Officer, Communicable Diseases Branch

Approval officer

Dr Sonya Bennett, Executive Director, Communicable Diseases Branch

Approval date: 10 June 2015

Review date: 10 June 2018

7. Version Control

Version	Date	Prepared by	Comments
1.0	2012	CHRISP	[QH-GDL-321-6-1:2012] Rescinded
2.0	March 2013	CHRISP	
3.0	January 2015	CDIM	