

---

## Radiation Safety Standard

---

### NOTIFICATION OF APPROVAL OF AN ALTERNATIVE IMAGE QUALITY TEST

I hereby approve the following test for Test 11 (Image Quality), of radiation safety standard HR001:2010 *Standard for radiation apparatus used to carry out plain film diagnostic radiography*.


#### Test

For extra-oral dental diagnostic radiation apparatus manufactured by Planmeca, the test specified in the document *Planmeca Digital X-ray Unit – Constancy test according to IEC 61223-3-4*, publication number 10008319 Version 2, published by Planmeca.

#### Criteria for Passing the Test

The quality of the image produced is within the limits specified in Planmeca's published acceptance test protocols for the X-ray unit being tested.

This approval commences on the 27 August 2010.

  
\_\_\_\_\_  
P M Vevers  
Delegate of the Chief Executive

27 August 2010  
Date